Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARWM-0027
Location:	LITTLE FALLS TOWNSHIP				
Calibration File No.:	00275	Calib. Date:	04/02/2014	Calib. No.:	00010
Certification File No.:	00276	Cert. Date:	04/02/2014	Cert. No.:	00008
Linearity File No.:	00277	Lin. Date:	04/02/2014	Lin. No.:	00008
Solution File No.:	00291	Soln. Date:	05/22/2014	Soln. No.:	00058
Sequential File No.:	00291	File Date:	05/22/2014		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXA S3-0015
Control Solution %:	0.100%			Expires:	11/05/2014
Solution Control Lot:	12K108			Bottle No.:	0696

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:17D		
Control 1 EC	0.099%	04:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:19D		
Control 2 EC	0.098%	04:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:20D		
Control 3 EC	0.099%	04:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PRESING First Name: JASON MI: D

Signature: #738

Badge No.: 738

Date: 05/22/2014