Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARWM-0027
Location:	LITTLE FALLS TOWNSHIP				
Calibration File No.:	00275	Calib. Date:	04/02/2014	Calib. No.:	00010
Certification File No.:	00276	Cert. Date:	04/02/2014	Cert. No.:	00008
Linearity File No.:	00277	Lin. Date:	04/02/2014	Lin. No.:	00008
Solution File No.:	00285	Soln. Date:	04/20/2014	Soln. No.:	00057
Sequential File No.:	00285	File Date:	04/20/2014		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXA S3-0015
Control Solution %:	0.100%			Expires:	11/05/2014
Solution Control Lot:	12K108			Bottle No.:	0697

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:26D		
Control 1 EC	0.099%	02:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:27D		
Control 2 EC	0.097%	02:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:29D		
Control 3 EC	0.098%	02:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PRESING First Name: JASON MI: -

Signature: #738 Badge No.: 738

Date: 04/20/2014