Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARWM-0027
Location:	LITTLE FALLS TOWNSHIP				
Calibration File No.:	00275	Calib. Date:	04/02/2014	Calib. No.:	00010
Certification File No.:	00276	Cert. Date:	04/02/2014	Cert. No.:	00008
Linearity File No.:	00277	Lin. Date:	04/02/2014	Lin. No.:	00008
Solution File No.:	00305	Soln. Date:	08/01/2014	Soln. No.:	00060
Sequential File No.:	00305	File Date:	08/01/2014		

Calibrating Unit:	WET	Model No.:	4	Serial No.:	DDXA S3-0015
Control Solution %:	0.100%			Expires:	01/06/2016
Solution Control Lot:	14A126			Bottle No.:	0743

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:59D		
Control 1 EC	0.100%	10:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01D		
Control 2 EC	0.099%	10:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D		
Control 3 EC	0.100%	10:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRIGGS First Name: JAMES MI: D

Signature: Det/SG James D Briggs

Badge No.: 734

Date: 08/01/2014