Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C		Serial No.: ARWM-0027
Location:	LITTLE FALLS TOWNSHIP		
Calibration File No.:	00275	Calib. Date: 04/02/2014	Calib. No.: 00010
Certification File No.:	00276	Cert. Date: 04/02/2014	Cert. No.: 00008
Linearity File No.:	00277	Lin. Date: 04/02/2014	Lin. No.: 00008
Solution File No.:	00314	Soln. Date: 09/05/2014	Soln. No.: 00061
Sequential File No.:	00314	File Date: 09/05/2014	

Calibrating Unit:	WET	Model No.: 4	Serial No.: DDXA S3-0015
Control Solution %:	0.100%		Expires: 01/06/2016
Solution Control Lot:	14A126		Bottle No.: 0735

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:06D		
Control 1 EC	0.101%	18:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:07D		
Control 2 EC	0.099%	18:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:08D		
Control 3 EC	0.100%	18:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PRALL First Name: BRYAN MI: M

Signature: _____ 740 Badge No.: 740

Date: 09/05/2014