Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARWM-0027
Location:	LITTLE FALLS TOWNSHIP				
Calibration File No.:	00316	Calib. Date:	10/02/2014	Calib. No.:	00011
Certification File No.:	00317	Cert. Date:	10/02/2014	Cert. No.:	00009
Linearity File No.:	00318	Lin. Date:	10/02/2014	Lin. No.:	00009
Solution File No.:	00323	Soln. Date:	11/03/2014	Soln. No.:	00063
Sequential File No.:	00323	File Date:	11/03/2014		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDZE-0030
Control Solution %:	0.100%			Expires:	01/06/2016
Solution Control Lot:	14A126			Bottle No.:	0739

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:03S		
Control 1 EC	0.100%	10:03S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04S		
Control 2 EC	0.098%	10:05S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06S		
Control 3 EC	0.098%	10:06S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRIGGS First Name: JAMES MI: D

Signature: DET/SGT

Badge No.: 734

Date: 11/03/2014